

Address: Building 2, Jinshui Rd, 107 Fudao, LGXC, Zhengzhou, 450018, China

Tel: +86 371 61317899

Fax: +86 371 86599698

Mail: info@plustrace.com

## Manufacturer's Certificate of Validation

Certificate No.	Product Name	Serial Number	Calibration Date
PL19121002	TEMPH	M19008107	Dec10,2019

### 1. Calibration Result of Temperature ( Unit: °C )

Calibration Point (°C)	Measured Value	Error	Result
-15.0	-14.8	0.2	PASS
15.0	15.1	0.1	PASS
30.0	30.1	0.1	PASS

### 2. Calibration Result of Relative Humidity ( Unit: %RH At 20°C )

Relative Value	Measured Value	Error	Result
40.00	40.9	0.9	PASS
60.00	62.2	2.2	PASS
80.00	82.0	2.0	PASS

PLUSTRACE TECHNOLOGY certifies that the products mentioned on the above pages has been thoroughly tested, validated and met performance accuracy specifications over the stated Ranges.

Approved By: *Peter Zhou*

Date: *2019/12/10*

河南省立光电子科技有限公司  
PLUSTRACE TECHNOLOGY CO., LTD